



HeartSine®

Lifesaving technology designed for the real world

Technical Note Ref:

Operation & Storage of HeartSine samaritan® PAD & Pad/Ped Paks at the correct Temperature.

Dear Valued Customer,

We would like to highlight the importance of storing and operating your HeartSine samaritan® PAD and associated Pad/Ped Paks at the correct temperatures as indicated below. Please note that this information can be found in the relevant User manuals.

For HeartSine samaritan® PAD devices with serial numbers beginning with 04, 05, 06, 07:

Standby Temperature:

The samaritan® PAD should be stored at a temperature of between 10 and 50°C (50 to 122°F).

HeartSine recommend that if possible the device is stored at a comfortable room temperature as this will ensure the peak performance for both the samaritan® PAD and the Pad-Pak.

Operation Temperature:

If the samaritan® PAD has been kept in accordance with the Standby Conditions (above), for a period of at least 24 hours, it may be used in responses where the ambient temperature is between 0 and 50°C (32 to 122°F). In this situation it may be safely used in the lower temperature range for a period of up to 1 hour after it has been exposed to the lower temperature.

Shipping and Transportation Temperature:

The samaritan® PAD may be temporarily stored in the range -10 to 50°C (14 to 122°F) for up to two days.

If you believe that the samaritan® PAD has been stored below 10°C (50°F) it should be returned to an ambient temperature of between 10 and 50°C (50 to 122°F) for a period of at least 24 hours before the device is considered ready for use.

Long term storage outside the temperature range 10 to 50°C may adversely affect the performance of the device.

Note: When considering location, avoid areas that expose the samaritan® PAD to extreme moisture, dust or temperatures.

We would also like to bring the following information relating to the HeartSine samaritan® PAD devices that have serial numbers starting with 04, 05, 06, 07 to your attention:

If during operation the samaritan® PAD delivers the voice prompt 'Warning Low Battery', and is suspected to have been subjected to a cold environment (below 10°C or 50°F), it may still be capable of operating correctly and safely; subject to the storage and operating temperature recommendations below:

Note: When the 'low battery' voice prompt is played, the green status indicator will stop flashing.

If the samaritan® PAD is returned to the recommended storage temperature range (10 to 50°C or 50 to 122 °F) when next switched on, either manually or by automatic self test, the green LED should resume flashing, therefore indicating the self-test to be completed successfully.

The warning is not related to low temperature if the status indicator does not start to flash green or the voice prompt 'Warning Low Battery' continues. In this instance contact your authorized HeartSine distributor or HeartSine Technologies' Technical Support directly.

If the samaritan® PAD is being used for an emergency response and 'Warning Low Battery' is heard, continue with CPR and defibrillation therapy with the device.

The samaritan® PAD may be capable of delivering typically up to ten defibrillation shocks when the prompt 'Warning Low Battery' is heard.

Note: While the samaritan® PAD continues to operate it will deliver safe and appropriate therapy.

If you have a spare Pad-Pak™ (battery and electrode cartridge) be prepared to replace the Pad-Pak™ battery.

In cold conditions ensure any spare Pad-Pak™'s are kept in accordance with recommendations stated above until required.



HeartSine®

Lifesaving technology designed for the real world

For HeartSine samaritan® PAD devices with serial numbers beginning with 08A & 09A:

Standby Temperature:

The samaritan® PAD should be stored at a temperature of between 0 and 50°C (32 to 122°F).

Heartsine recommend that if possible the device is stored at a comfortable room temperature as this will ensure the peak performance for both the samaritan® PAD and the Pad-Pak™.

Operation Temperature:

If the samaritan® PAD has been kept in accordance with the Standby Conditions (above), for a period of at least 24 hours, it may be used in responses where the ambient temperature is between 0 and 50°C (32 to 122°F). In this situation it may be safely used in the lower temperature range for a period of up to 1 hour after it has been exposed to the lower temperature.

Shipping and Transportation Temperature:

The samaritan® PAD may be temporarily stored in the range -10 to 50°C (14 to 122°F) for up to two days.

If you believe that the samaritan® PAD has been stored below 0°C (32°F) it should be returned to an ambient temperature of between 0 and 50°C (32 to 122°F) for a period of at least 24 hours before the device is considered ready for use.

Long term storage outside the temperature range 0 to 50°C (32 to 122 °F) may adversely affect the performance of the device.

Note: When considering location, avoid areas that expose the samaritan® PAD to extreme moisture, dust or temperatures.

We would also like to bring the following information relating to the HeartSine samaritan® PAD devices that have serial numbers starting with 08A or 09A to your attention:

If during operation the samaritan® PAD delivers the voice prompt 'Warning Low Battery', and is suspected to have been subjected to a cold environment (below 0°C or 32°F), it may still be capable of operating correctly and safely; subject to the storage and operating temperature recommendations below:

Note: When the 'low battery' voice prompt is played the status indicator will begin to flash red.

If the samaritan® PAD is returned to the recommended storage temperature range (0 to 50°C or 32 to 122 °F) when next switched on, either manually or by automatic self test, the flashing red LED should have changed to green, therefore indicating the self-test to be completed successfully.

The warning is not related to low temperature if the status indicator continues to flash red or the voice prompt 'Warning Low Battery' continues. In this instance contact your authorized HeartSine distributor or HeartSine Technologies' Technical Support directly.

If the samaritan® PAD is being used for an emergency response and 'Warning Low Battery' is heard, continue with CPR and defibrillation therapy with the device.

The samaritan® PAD may be capable of delivering typically up to ten defibrillation shocks when the prompt 'Warning Low Battery' is heard.

Note: While the samaritan® PAD continues to operate it will deliver safe and appropriate therapy.

If you have a spare Pad-Pak™ (battery and electrode cartridge) be prepared to replace the Pad-Pak™ battery.

In cold conditions ensure any spare Pad-Pak™'s are kept in line with recommendations stated above until required.

H009-013-005-0